




CERTIFICATE

No. U8 09 02 69406 001

Holder of Certificate: Charles River Laboratories, Inc.
 251, Ballardvale Street
 Wilmington MA 01887-1096
 USA

Production Facility(ies): 69137

Certification Mark:



Product: Analysis Equipment
 (Portable Test System)

Model(s): Endosafe PTS

Parameters:

Rated voltage:	12 Vdc
Rated current:	5 A
Protection class:	III

Tested according to: UL 61010-1:2004
 CAN/CSA C22.2 No. 61010-1:2004

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. See also notes overleaf.

Test report no.: 028-71323158-100

Date, 2009-02-16

Page 1 of 1

